



## DECLARATION OF CONFORMITY

Product Identification:

First declared: 2006

**Fluke 561 - IR Thermometer**

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO 9001:2015 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

Directive 2011/65/EU + 2015/863/EU, UK Regulation S.I. 2012:3032 - Reduction of Hazardous Substances (RoHS)  
Directive 2014/30/EU, UK Regulation S.I. 2016:1091 - Electromagnetic Compatibility (EMC)  
Regulation 2023/1542/EC, UK Regulation S.I. 2009:890+2015:63 - Batteries

Standards Used:

EN/BS 63000:2018 Assessment of electrical and electronic products for restriction of hazardous substances  
EN/BS 61326-1:2013 Electrical equipment for measurement, control and laboratory-EMC; General  
EN/BS 61326-2-2:2013 Particular EMC requirements for portable test, measuring and monitoring equipment  
EN/BS 61010-1:2010+A1:2019 Safety requirements for measurement, control and laboratory equipment; General  
EN/BS 60825-1:2014 Safety of laser products; Classification and requirements

Special Conditions:

Not intended for hand-held use at voltages above 30 Vrms, 60 Vdc or 42.4 Vpeak.

EU Importer/Distributor

Fluke Europe B.V.  
PO Box 1186  
5602 BD Eindhoven  
The Netherlands

UK Importer/Distributor

Fluke Precision Measurement Ltd.  
52 Hurricane Way  
Norwich, NR6 6JB  
United Kingdom

Manufacturer

Fluke Corporation  
6920 Seaway Boulevard M/S 143F  
Everett, WA 98203; USA

Henry Kuster  
EMEA Logistics and Compliance Manager  
Fluke Europe B.V.

Thomas M. Smith P.E.  
Product Compliance Technical Manager  
Fluke Corporation